

Clinical examination of the eyes functional status and assessment of equivalent dose to eye lens in medical staff performing endovascular interventions under X-ray guidance

Ryzhkin S., Slesareva A., Galeeva G., Ivanov S.
Kazan Federal University, 420008, Kremlevskaya 18, Kazan, Russia

Abstract

Growing use of X-ray guided surgery interventions leads to increase in the risk of radiation side effects in medical staff participating in image-guided interventional procedures. In this connection to protect health and ensure professional longevity of the interventionists are important scientific and social problems. The purpose of the study presented in the article is to assess radiation dose to the lens of eyes, doses to whole body and hands of medical staff performing X-ray-guided endovascular procedures. Doses were measured with the use of thermoluminescent dosimeters. Due to clinical eye examination for above staff to assess abilities to refract the light existence of involution changes, which are unusual for a specific age of an examined person, are found. The study proves that new annual equivalent dose limit for the lens of the eyes of 20 mSv recommended by the IAEA is justified.

<http://dx.doi.org/10.21870/0131-3878-2017-26-3-90-99>

Keywords

Biomicroscopy of eye, Endovascular intervention, Equivalent dose, Lens of the eye, Radiation safety of medical staff, Radiation-associated cataract, Thermoluminescence dosimetry, X-ray image

References

- [1] Kotenko K.V., Bushmanov A.Yu., Tyurin I.Ye., Kostylev V.A., Tkachev S.I., Dolgushin B.I., Boyko A.A., Narkevich B.Ya. On the issue of hazardous conditions in the radiology departments of medical institutions. *Meditsinskaya radiologiya i radiatsionnaya bezopasnost'*-Medical Radiology and Radiation Safety, 2013, vol. 58, no. 4, pp.17-22. (In Russian).
- [2] Fish D.E., Kim A., Ornelas C., Song S., Pangarkar S. The risk of radiation exposure to the eyes of the interventional pain physician. *Radiol. Res. Pract.*, 2011, vol. 2011, pp. 1-5. doi: 10.1155/2011/609537.
- [3] Tsyb A.F., Abakushina Ye.V., Abakushin D.N., Romanko Yu.S. Ionizing radiation as a risk factor for the development of radiation cataracts. *Mediko-biologicheskie problemy zhiznedeyatel'nosti*-Medical-biological Problems of Life, 2013, no. 1 (9), pp. 34-41. (In Russian).
- [4] Basic anatomical and physiological data for use in radiological protection: Reference values: ICRP Publication 89. *Annals of the ICRP*, 2002, vol. 32, no. 3-4, pp. 1-277.
- [5] Ostrovskiy M.A. Light and radiation ecology of vision. *Vestnik Rossiyskoy voenno-meditsinskoy akademii*-Bulletin of the Russian Military Medical Academy, 2008, vol. 23, no. 3, pp. 9-12. (In Russian).

- [6] Radiation safety standard (NRB-99/2009): Sanitary rules and norms (SanPiN 2.6.1.2523-09): appr. and put into effect on 07.07.09. Moscow, Federal Center of Hygiene and Epidemiology, 2009. 100 p. (In Russian).
- [7] Galeyeva G.Z., Ryzhkin S.A., Sergeyeva S.Yu. Effects of ionizing radiation on the human body and the eyes. *Prakticheskaya meditsina-Practical Medicine*, 2016, no. 7(99), pp. 37-41. (In Russian).
- [8] Ivanov S.I., Loginova S.V., Akopova N.A., Okhrimenko S.Ye., Nurlybayev K.N. Problems of eye lens dosimetry. *Meditinskaya radiologiya i radiatsionnaya bezopasnost'-Medical Radiology and Radiation Safety*, 2014, vol. 59, no. 4, pp. 67-72. (In Russian).
- [9] Rubtsov V.I., Klochkov V.N., Trebukhin A.B., Nefedov A.Yu., Klochkova Ye.V., Nurlybayev K.N., Martynyuk Yu.N., Medvedev L.L. Monitoring of dose equivalent to the lens of the eye and evaluation of the possibility of its reduction. *ANRI-ANRI*, 2013, no. 3, pp. 32-37. (In Russian).
- [10] IAEA. Radiation Protection and Safety of Radiation Sources: International Basic Safety Standards, ? GSR Part 3. Vienna, IAEA, 2015. 482 p.